



Italian  
Medical  
Touch

CA-MI snc  
Via Ugo La Malfa N°31  
43010 Pilastro (PR) Italia  
Tel. 0521 / 631138  
Fax. 0521 / 639041  
Cod. Fisc. / P.IVA 00977090349  
www.ca-mi.it

**ISO 9001 – ISO 13485 CERTIFIED**



**Certificate ISO 9001 no. 800210**  
**Certificate ISO 13485 no. 908016**

## DECLARATION OF CONFORMITY

**Manufacturer:** CA-MI snc

**Address:** Via Ugo La Malfa n°31

**Device:** Incentive Spirometers

**Trade name, model:** PULMOLIFT – REF RE 400250

**Intended purpose:** Incentive Spirometers to exercise and stimulate respiration through expiration. Can be used alone or connected to any other CA-MI Incentive Spirometers -such as Pulmogain or Pulmovol- to exercise respiration through a complete cycle of inspiration and expiration without interruptions.

With the present one attests that such device is in compliance with the MDD 93/42/CEE .

The mentioned device has been classified Medical Device with risk **I class the not sterile** in agreement to the rules of Attached the IX of the MDD 93/42/CEE.

EC marking procedure has been carried out according to the provisions of Annex VII (Declaration of Conformity)

**Date of issued:** lunedì 13 novembre 2006

**Issued by**  
Quality Assurance Manager  
Manuel Saccani

**Verified by:**  
General Manager  
Mario Attolini